

Notice of Allowability

Application No.

10/009,631

Examiner

George Manuel

Applicant(s)

HUNYOR ET AL.

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☐ This communication is responsive to _____.
2. ☒ The allowed claim(s) is/are 1-7 and 14-37.
3. ☒ The drawings filed on 12 December 2001 are accepted by the Examiner.
4. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☒ All b) ☐ Some* c) ☐ None of the:
 1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).


* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☒ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____


George Manuel
Primary Examiner
Art Unit: 3762

DETAILED ACTION

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it **MUST** be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Doug Price on 1/18/05.

The application has been amended as follows:

In claim 1, line7, after "to" insert, --pull at least that region of the heart in a direction relatively toward the distal wall and thereby--.

Allowable Subject Matter

Claims 1-7 and 14-37 are allowed.

The following is an examiner's statement of reasons for allowance:

The claimed features of: (a) pulling at least a region of the heart to decompress the heart, (b) the heart compressing wall and the distal wall of a main body being of materials with different degrees of stiffness, the distal wall including a reinforcing material therein to provide for a greater degree of stiffness relative to the heart compressing wall, and (c) at least a portion of the heart compressing wall including a biointegratable material surface which facilitates the ingrowth of vascularised cellular tissue elements into the wall, the ingrowth of tissue into the heart compressing surface serving to affix the heart compressing wall of the main body to the heart as claimed are not taught nor suggested by the prior art of record.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably

Art Unit: 3762


accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Zadini et al '364 discloses a cardiac pump with an expandable balloon 105 which is capable of compressing the heart and decompressing the heart. There lacks a teaching or suggestion and it does not appear obvious to modify the device of Zadini et al to pull the heart tissue to decompress the heart.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George Manuel whose telephone number is (571) 272-4952.


George Manuel
Primary Examiner
Art Unit: 3762

1/18/05